

Tip-Sample deconvolution problems

Surfaces A, B and C are measured as the resulting *Measured topography* graph due to the tip geometry. In all these cases the result is exactly the same so the reverse procedure (tip-sample deconvolution) is not deterministic.

